



MEETING ABSTRACT

Open Access

A prototype RT-PCR assay for detection of XMRV in multiple human sample types

Ning Tang^{1*}, Andrea Frank¹, Robert Kowal¹, Gregor Leckie¹, John Hackett Jr², Graham Simmons³, Michael Busch³, Klara Abravaya¹

From 15th International Conference on Human Retroviruses: HTLV and Related Viruses
Leuven and Gembloux, Belgium. 5-8 June 2011

Background

Xenotropic murine leukemia virus-related virus (XMRV) has been reported to be associated with prostate cancer and chronic fatigue syndrome. To help resolve the role of XMRV in human disease, it is critical to develop sensitive and accurate PCR assays for XMRV detection.

Materials and methods

The automated m2000 system RT-PCR assays detect the pol and env regions of XMRV in whole blood, plasma, urine, and cell pellets. Assay performance was assessed by testing two blinded panels prepared by the Blood XMRV Scientific Research Working Group (SRWG), as well as clinical specimens.

Results

For the SRWG whole blood panel, all six XMRV negative samples were assay negative, while all the 22Rv1-spiked samples from 0.5 cells/mL to 9,900 cells/mL were detected (3/3 for each panel member). For the SRWG plasma panel, all six XMRV negative samples were assay negative, while the 22Rv1 supernatant-spiked plasma samples were detected as positive, 1/3 for the 3.2 XMRV copies/mL panel member, 2/3 for the 16 copies/mL panel member, and 3/3 for panel members containing > 80 copies/mL. For screening of clinical specimen, assay positive results were obtained for 10% (2/20) of prostate cancer FFPE specimens, 0.5% (2/400) of prostate cancer urine pellets, and 0% (0/135) of the cervical swab specimens.

Conclusions

We developed an automated high throughput real-time RT-PCR assay (with internal control) to detect XMRV.

Use of this assay should assist in elucidation of the role of XMRV in human disease.

Author details

¹Abbott Molecular Inc., Des Plaines, IL, 60018, USA. ²Abbott Diagnostics, Abbott Park, North Chicago, Illinois, 60064, USA. ³Blood Systems Research Institute, University of California, San Francisco, CA, 94118, USA.

Published: 6 June 2011

doi:10.1186/1742-4690-8-S1-A220

Cite this article as: Tang et al.: A prototype RT-PCR assay for detection of XMRV in multiple human sample types. *Retrovirology* 2011 **8**(Suppl 1):A220.

Submit your next manuscript to BioMed Central and take full advantage of:

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at
www.biomedcentral.com/submit



* Correspondence: Ning.Tang@abbott.com

¹Abbott Molecular Inc., Des Plaines, IL, 60018, USA

Full list of author information is available at the end of the article